

ECG, EEG, Breathing Signals, and Machine Learning: Computer-Aided Detection of Obstructive Sleep Apnea Syndrome and Depression

Mostafa M Moussa¹, Yahya Alzaabi¹, Ahsan Khandoker¹

¹ Khalifa University, Abu Dhabi, UAE

Abstract

Obstructive Sleep Apnea Syndrome (OSAS) and Major Depressive Disorder (MDD) are both common conditions associated with poor quality of life. We seek to classify OSAS and depression in OSAS patients, as well as sleep stages using multiple machine learning techniques. We have extracted features from 5-minute intervals of electrocardiograms (ECG), breathing signals, and electroencephalograms (EEG) recorded from a total of 118 subjects, of which 89 are used for training and 10-fold cross-validation and 29 are used for testing or a 75/25% split. The best classification performance of OSAS was obtained with light sleep and deep sleep with ReliefF using random forest and boosted trees, respectively. It has yielded an accuracy of 100.00%, F1-Score of 100.00%, Cohen's κ Coefficient of 1.00, and a Matthews correlation coefficient (MCC) of 1.00. All sleep stages with 10 principal components using random forest yielded an accuracy of 77.50%, F1-Score of 78.05%, Cohen's κ of 0.571, and an MCC of 0.632 for classification of depression in OSAS patients. Sleep staging was best done using bagged trees with features selected via sequential backward feature selection, yielding an accuracy of 76.90%, F1-Score of 75.90%, Cohen's κ of 0.480, and an MCC of 0.634. These results show promise in detecting OSAS and depression in OSAS patients, particularly using light and deep sleep data.

1. Introduction

Obstructive Sleep Apnea Syndrome (OSAS) is a prevalent condition with 3-7% of men and 2-5% of women in the general population [1] suffering from it with an associated decrease in sleep quality and overall quality of life, due to its many comorbidities. These include strokes, coronary heart disease (CHD), mood disorders, memory loss, and reduced cognitive performance [2–4]. Electrocardiography (ECG) is among the most common electrophysiological signal recording techniques due to the importance of reading heart signals for countless applications and tests, and as such, it is also commonly recorded in sleep studies,

or polysomnography (PSG), often for use with machine learning algorithms.

Works like Khandoker *et al.*'s, Bozkurt *et al.*'s and Erdenebayar *et al.*'s focus on the use of support vector machines (SVMs), multiple classic classifiers, and convolutional and recurrent neural networks (CNNs and RNNs), respectively [4–6]. The first of which obtained a testing accuracy of 92.85% and a Cohen's κ coefficient of 0.85 with 83 out of 125 subjects after training SVMs with heart rate variability (HRV) and ECG-derived respiration (EDR) features and leave-one-out cross-validation [4]. Bozkurt *et al.* derived HRV and QRS intervals from 10 ECGs, and in turn, extracted features and performed feature selection using Fisher's algorithm. An ensemble of decision trees (DT), Kth-nearest neighbor (KNN), and SVM and 50% of the features selected, yielded the best performance with an accuracy of 85.12%, sensitivity of 85.00% and specificity of 86.00% [5]. Finally, Erdenebayar *et al.* use gated recurrent unit (GRU) neural networks, which are a type of long short-term memory (LSTM) recurrent neural networks (RNN) with 80% of 86 subjects used for training and the remainder for testing, to classification accuracy and sensitivity of 99.00% [6].

Works regarding the detection of OSAS and depression are somewhat rare, though machine learning is a relatively standard methodology for detecting these individual conditions. Furthermore, the discussed works commonly extensively pre-process data in order to have robust features for machine learning. Hence, the contribution of this work centers around the idea of using features extracted from 5-minute intervals from the ECG, EEG, and breathing signals following filtering and re-sampling as necessary, selecting features among these using various algorithms. Afterward, we use these features to train machine learning algorithms to perform sleep staging, classify OSAS, and classify depression in OSAS patients.

We describe the methodology utilized in Section 2 and the results obtained from following said methodology in Section 3, we discuss our results in Section 4. Finally, we end with concluding remarks and possible avenues for future work in Section 5.

2. Methodology

Overnight polysomnography was performed on 86 subjects at the American Center for Psychiatry and Neurology, among which 40 had OSAS alone, 40 had OSAS paired with depression. In addition, 32 subjects without OSAS or depression are taken from the Stanford Technology Analytics and Genomics in Sleep (STAGES) dataset [7] to form a total of 118 subject-dataset with 59 male and 57 female subjects, of which 89 or 75% are used for training and 10-fold cross-validation and the remaining 29 or 25% are used for testing.

We mainly focus on ECG, EEG, and breathing signals (flow and thorax signals), the first two were sampled at 100 Hz and 200 Hz, respectively, and were re-sampled to breathing signals' sampling frequency, 10 Hz, to synchronize the signals in time prior to feature extraction. Power-line interference is filtered out of the electrophysiological signals via a 50 Hz notch filter, followed by band-pass filtering as appropriate before re-sampling. Finally, the signals are split into 5-minute intervals, from which features are extracted based on the status of depression, apnea, and sleep stage during that interval, with depression being dependent on the subject themselves, rather than the interval.

Classification of OSAS and classification of depression in OSAS patients are done both with all data, and the data for each individual sleep stage, while sleep staging is logically only done with all data. Since we begin by classifying whether or not the subjects have OSAS using subjects with OSAS regardless of depression status vs the control group, the control group is removed afterward when we seek to classify depression. This means that the number of intervals or observations available for the classification of depression is less than the number of observations available for detection of OSAS and sleep staging (with all data only). Consequently, the selected features for the classification of depression would also differ from those selected for the classification of OSAS and for sleep staging.

The features extracted include respiratory frequency extracted from breathing signals, and brain wave powers extracted from EEG channels O1, O2, C3, C4, F3, and F4 with A1 and A2 (or M1 and M2) as reference channels. Furthermore, very low-frequency, low-frequency, and high-frequency powers, as well as the ratio of low-to-high-frequency powers of ECG signals or heart rate variability signals, and features pertaining to both ECG and breathing were extracted such as lambda, which represents the phase coupling between R-R intervals in the ECG and thorax signal, and respiratory sinus arrhythmia (RSA), which represents the reduction in R-R interval duration during inspiration and extension during expiration. The resultant features are thus lambda, RSA, respiratory frequency, normalized RSA, vLF power, LF power, normalized LF power, HF power, normalized HF power, LF/HF,

and the band powers of beta, theta, alpha, and delta from each of the 6 EEG channels aside from the reference ones for a total of 34 features.

No further processing is applied to this data, but features are selected using one of 7 techniques/algorithms: sequential forward feature selection (SFFS), sequential backward feature selection (SBFS), minimum redundancy maximum relevance (MRMR), ReliefF [8], neighborhood component analysis (NCA), Chi-Squared, principal component analysis taking the first 10 components, and principal component analysis taking components that explain 95% of the variance.

After feature selection, the data is ready to be input into our classifiers. We essentially have two sets of problems, the first set involves three problems: sleep staging, done in parallel with the classification of OSAS, and classification of depression in OSAS patients. The second set involves the latter two problems performed with data from each individual sleep stage. This means we have, overall, three classification problems, the first of which is 3-class due to the way sleep stages were denoted (light sleep, REM sleep, and deep sleep were considered; light sleep is not split into two stages), and the second and third are binary. We end up using up to 14 classifiers including Gaussian Distribution Naive Bayes (NB), linear discriminant analysis (LDA) with a regularization term γ of 0, decision tree (DC) with 100 maximum number of branch nodes split by Gini's diversity index (GDI), Euclidean distance k^{th} nearest neighbor (KNN) with k set to 1, radial basis function (RBF) support vector machine (SVM) with automatically computed kernel scale and a box constraint of 1, bagged trees with 1290 maximum number of branch nodes and 30 learning cycles, random forest (RF), boosted trees with 4 algorithms (AdaBoost, RUSBoost, LogitBoost, GentleBoost) and a maximum number of branch nodes of 30, subspace KNN and discriminant with 30 learning cycles, and a 4-layer artificial neural network (ANN), wherein there are an input and output layers and two hidden layers with 100 units in each.

3. Results

In order to measure the performance of our classifiers, we use several well-known metrics in machine learning, namely accuracy, sensitivity (also known as recall), specificity, precision, F1-score, Cohen's Kappa (κ) coefficient, Matthews correlation coefficient (MCC), and the area under the receiver operating characteristics (ROC) curve, also known as AUC. These metrics are described in Equations 1 through 7, where TP, FP, FN, and TN are true positive, false positive, false negative, and true negative instances and the expected accuracy is the accuracy when the instances are classified by chance, based on the number of classes and number of instances in each class.

$$Accuracy = \frac{TP + TN}{TP + FP + FN + TN} \quad (1)$$

$$Sensitivity = \frac{TP}{TP + FN} \quad (2)$$

$$Specificity = \frac{TN}{FP + TN} \quad (3)$$

$$Precision = \frac{TP}{TP + FP} \quad (4)$$

$$F1 - Score = \frac{2 * Precision * Sensitivity}{Precision + Sensitivity} \quad (5)$$

$$\kappa = \frac{Accuracy - ExpectedAccuracy}{1 - ExpectedAccuracy} \quad (6)$$

$$MCC = \frac{(TP * TN) - (FP * FN)}{\sqrt{(TP + FP)(TP + FN)(TN + FP)(TN + FN)}} \quad (7)$$

Sleep Staging is best done with SBFS using bagged trees for classification with an AUC of 0.846, an accuracy of 76.90%, sensitivity of 75.38%, specificity of 87.04%, precision of 77.03%, F1-score of 75.90%, Cohen’s κ of 0.480, and a Matthews correlation coefficient of 0.634. OSAS is best classified with ReliefF using light and deep sleep data, using random forest and boosted trees with GentleBoost algorithm, respectively for classification. These algorithms with their respective data and feature selection techniques yielded an AUC, Cohen’s κ , and Matthews correlation coefficient of 1.00, as well as accuracy, sensitivity, specificity, precision, and F1-score of 100%. Depression in OSAS patients is best classified with PCA taking the first 10 PCs using deep sleep data, using a random forest for classification with an AUC of 0.880, an accuracy of 77.50%, sensitivity of 64.00%, specificity of 100.00%, precision of 100.00%, F1-score of 78.05%, Cohen’s κ of 0.571, and a Matthews correlation coefficient of 0.632. These results are summed up in Table 1, and Figure 1 provides a visual representation in the form of posterior probability and box plots.

Table 1. Best classification results for each problem.

	Sleep Staging	Classification of OSAS		Classification of Depression in OSAS Patients
		Light Sleep	Deep Sleep	Deep Sleep
Sleep Stage	N/A or All Stages			
Feature Selection Technique	SBFS	ReliefF	ReliefF	PCA-10 PCs
Model	Bagged Trees	Random Forest	GentleBoost Boosted Trees	Random Forest
AUC	0.846	1.00	1.00	0.880
Accuracy (%)	76.90	100.00	100.00	77.50
Sensitivity (%)	75.38	100.00	100.00	64.00
Specificity (%)	87.04	100.00	100.00	100.00
Precision (%)	77.03	100.00	100.00	100.00
F1-Score (%)	75.90	100.00	100.00	78.05
κ	0.480	1.00	1.00	0.571
MCC	0.634	1.00	1.00	0.632

4. Discussion

The computed metrics in Table 1 show excellent performance for classification of OSAS with these datasets and methodology, as well as decent performance for sleep staging and classification of depression in OSAS patients. Deep sleep seems to be the common denominator in the best classification performance for both OSAS and depression in OSAS patients, allowing us to recommend recording polysomnography signals during deep sleep. Since our classification of OSAS and depression using all sleep stage data was done in parallel with sleep staging, we refrain from making hard recommendations for automation of sleep staging before moving on to classification of OSAS and depression since its classification performance can likely be improved with further processing of the signals or more advanced classification techniques. Moreover, note that the performance using REM sleep data for classification does not appear in our table, which indicates that increased brain activity may be a slight hindrance for all three classification problems when using ECG, EEG, and breathing signals, perhaps because 24 out of the initial 34 features are extracted from the EEG. In addition, the results obtained for OSAS and depression classification using all sleep stage data compared to individual sleep stages is slightly worse, so we recommend clinically or mathematically defining sleep stages prior to classification of OSAS and depression in OSAS patients. We also notice that decision tree ensemble classifiers, namely bagged trees, boosted trees, and random forest yield the best performance in all three classification problems, meaning we could forgo other ML algorithms.

Future directions in this research should include performing classification using deep learning with the PSG signals directly without manual feature extraction, which allows better automated analysis of PSG data, or simply further processing of the signals or features. Additionally, interpretability measures like Shapley additive explanations (SHAP) can be computed to more accurately analyze feature importance.

5. Conclusion

To wrap up, the main goal of our work is to classify OSAS, and depression in OSAS patients using 14 different machine learning and deep learning techniques from ECG, EEG, and breathing signals extracted from overnight polysomnography (PSG) recorded 118 subjects, using the data of 89 for training and 10-fold cross-validation and 29 for testing. Best sleep staging performance was obtained using bagged trees with features selected using sequential backward forward selection, with an accuracy of 76.90%, an F1-score of 75.90%, a κ of 0.480 and a MCC of 0.634. However, intervals of deep sleep yielded the best results for




Figure 1. A) Posterior probability of the best staging configuration. B) Posterior probability of one of the two best OSAS classification configurations (deep sleep, reliefF, GentleBoost). C) Posterior probability of the best depression with OSAS classification configuration.

classification of OSAS and depression in OSAS patients, with an accuracy of 100.0%, an F1-score of 100.0%, a κ of 1.00, and an MCC of 1.00 with features selected using the ReliefF algorithm for OSAS using boosted trees with the GentleBoost algorithm, and principal component analysis taking the first 10 principal components and using random forest to classify depression in OSAS patients yielded an accuracy of 77.50%, an F1-score of 78.05%, a κ of 0.571 and an MCC of 0.632, respectively. Light sleep data with ReliefF and random forest for OSAS classification also yielded the same performance.

Acknowledgement

The authors would like to thank the American Center for Psychiatry and Neurology (ACPN) in Abu Dhabi for their invaluable contribution in sharing the polysomnography data, and acknowledge the support of the biomedical engineering department and the Healthcare Engineering Innovation Center (HEIC) at Khalifa University of Science and Technology. The authors would also like to highlight the importance of the KAU-KU Joint Research Program, in particular, project DENTAPNEA between Khalifa University (Project number 8474000376) and King Abdulaziz University, in particular the advice of Dr. Angari, Dr. Balamesh, Dr. Khraibi, and Dr. Marghalani.

References

- [1] Young T, Palta M, Dempsey J, Skatrud J, Weber S, Badr S. The occurrence of sleep-disordered breathing among middle-aged adults. *New England Journal of Medicine* 1993; 328(17):1230–1235.
- [2] Mannarino MR, Di Filippo F, Pirro M. Obstructive sleep apnea syndrome. *European Journal of Internal Medicine* 2012; 23(7):586–593.

- [3] Li M, Li X, Lu Y. Obstructive sleep apnea syndrome and metabolic diseases. *Endocrinology* 2018;159(7):2670–2675.
- [4] Khandoker AH, Palaniswami M, Karmakar CK. Support vector machines for automated recognition of obstructive sleep apnea syndrome from ECG recordings. *IEEE Transactions on Information Technology in Biomedicine* 2008; 13(1):37–48.
- [5] Bozkurt F, Uçar MK, Bozkurt MR, Bilgin C. Detection of abnormal respiratory events with single channel ECG and hybrid machine learning model in patients with obstructive sleep apnea. *IRBM* 2020;41(5):241–251.
- [6] Erdenebayar U, Kim YJ, Park JU, Joo EY, Lee KJ. Deep learning approaches for automatic detection of sleep apnea events from an electrocardiogram. *Computer Methods and Programs in Biomedicine* 2019;180:105001.
- [7] Zhang GQ, Cui L, Mueller R, Tao S, Kim M, Rueschman M, Mariani S, Mobley D, Redline S. The national sleep research resource: towards a sleep data commons. *Journal of the American Medical Informatics Association* 2018; 25(10):1351–1358.
- [8] Kononenko I, Šimec E, Robnik-Šikonja M. Overcoming the myopia of inductive learning algorithms with reliefF. *Applied Intelligence* 1997;7(1):39–55.

Address for correspondence:

Mostafa Moussa
 Department of Biomedical Engineering, Khalifa University of Science and Technology, Abu Dhabi, UAE, 127788
 mostafa.moussa@ku.ac.ae